

KU/PMU-D/D IV/6873/2025

24.06.2025



**KANNUR UNIVERSITY**

**Re-Quotation Notice**

**Re-Quotation for the Supply and Installation of Peristaltic Pump with Accessories at, Dept. of Molecular Biology, Kannur University, Dr. Janaki Ammal Campus, Palayad, Thalassery.**

Sealed competitive Quotations are invited from the authorized dealers having proven track record for the "Supply and Installation of Peristaltic Pump with Accessories" at Department of Molecular Biology, Dr. Janaki Ammal Campus, Palayad under PM USHA Scheme. Quotation with details of model, price etc. should reach the Registrar, Kannur University, Civil Station P.O, Kannur -670002 on or before **10/07/2025, 2 PM** and late quotations will not be accepted. The quotations will be opened at **3 PM** on the same day in the presence of such of the quotationers or their authorized representatives who may be present at that time.

**SCHEDULE**

|  |   |
|--|---|
| QUOTATION NOTICE NO                    | KU/PMU-D/D IV/6873/2025   |
| NAME OF WORK                           | Supply and Installation of Peristaltic Pump with Accessories      |
| LOCATION TO SUPPLY & INSTALL           | Department of Molecular Biology, Dr. Janaki Ammal Campus, Palayad |
| QUOTATION SUBMISSION END DATE AND TIME | <b>10/07/2025, 2 PM</b>   |
| QUOTATION OPENING DATE AND TIME        | <b>10/07/2025, 3 PM</b>   |

**PRODUCT INFORMATION AND SPECIFICATION**

| SI No. | Product Specification  | Required Qty |
|--------|--|--------------|
| 1.     | Flow Range: 0.5ml/hr/ch to 2500ml/hr/ch<br>No of Step: 1400<br>No of Channels: 2<br>No of Rollers: 6<br>Tube Size: ID:2mm to 5mm WT:1.5mm<br>Display: LED Display<br>Programmable timer from 1 min to 24 Hours<br>Non Volatile Memory backup /Audible Alarm with Auto stop<br>Includes: Silicone Rubber Tubing 5mm ID X 8mm OD -5 meters.<br>Accuracy +/- 2% (Reproducibility) | 1            |

|   |  |
|---|--|
| RS232 computer interface<br>Power 230 V AC, 50 Hz<br>Warranty: 1 Year |  |
|---|--|

**TERMS AND CONDITIONS**

1. The total amount quoted should be inclusive of GST and all other installation charges, if any. The base rate, GST charges and installation charges should be furnished separately.
2. Quotation should be valid for a minimum period of **three months**.
3. Details of Warranty Condition and details of nearest service/maintenance centre should be provided along with quotation.
4. Supplier shall deliver and Install the item at the locations mentioned above.
5. The Registrar, Kannur University reserves the right to accept or reject the quotation and to trade or not to trade the old stores, without assigning any reason.

*GSTIN 32AAAGK0152J1ZT*

**Contact: 04972715243**

*Copy to : Web Manager/Notice Board*

**Sd/-**  
**REGISTRAR**